
Colorimetric Determination Of An Equilibrium Constant In Aqueous Solution

THE THEORETICAL FOUNDATION:
The Colorimetric Determination Of An Equilibrium Constant In Aqueous Solution
Downloaded from db.whlarchitecture.com
by Alvarado & Good


INTRODUCTION TO COLORIMETRIC EQUILIBRIUM ANALYSIS

Understanding chemical equilibrium is a cornerstone of physical chemistry, and the ability to quantify equilibrium constants in solution is essential for fields ranging from biochemistry to environmental science. Among the most accessible and elegant methods for determining an equilibrium constant in aqueous solution is colorimetry. This approach leverages the simple relationship between the concentration of a colored species and the intensity of light it absorbs. By measuring how much light a solution absorbs at a specific wavelength, we can deduce concentrations and, through careful stoichiometric reasoning, calculate the equilibrium constant for a reversible reaction. This article provides a comprehensive walkthrough of the theory, experimental procedure, data analysis, and practical applications of the colorimetric determination of an equilibrium constant in aqueous solution, using the classic reaction between iron(III) ions and thiocyanate ions as a guiding example.

CHEMICAL EQUILIBRIUM AND COLORIMETRY

Chemical Equilibrium and the Equilibrium Constant

In a reversible chemical reaction, equilibrium is reached when the rates of the forward and reverse reactions become equal, and the concentrations of reactants and products remain constant over time. For a general reaction in aqueous solution:



the equilibrium constant K_c (expressed in terms of molar concentrations) is defined as:

$$K_c = \frac{[C]^c [D]^d}{[A]^a [B]^b}$$

The value of K_c is constant at a given temperature and provides quantitative insight into the position of equilibrium. A large K_c indicates that products are favored at equilibrium, while a small K_c indicates reactants are favored. In the context of colorimetric determination, we must be able to measure at least one

species' concentration directly at equilibrium, and from that, calculate all others using initial concentrations and the reaction stoichiometry.

Principles of Colorimetry and the Beer-Lambert Law

Colorimetry is based on the principle that many substances—particularly transition metal complexes and organic dyes—absorb visible light. The amount of light absorbed is proportional to the concentration of the absorbing species and the path length of the light through the solution. This relationship is formalized by the Beer-Lambert Law:

$$A = \epsilon b c$$


Where:

- **A** = absorbance (dimensionless)
- **ϵ** = molar absorptivity ($\text{L mol}^{-1} \text{cm}^{-1}$), a constant for a given substance at a specific wavelength
- **b** = path length of the cuvette (usually 1 cm)
- **c** = concentration of the absorbing species (mol L^{-1})

For the colorimetric determination of an equilibrium constant, we first need to establish a calibration curve: a plot of absorbance versus known concentrations of the colored species. Then, at equilibrium, we measure the absorbance of the solution and use the calibration curve to find the equilibrium concentration of that species. This concentration is the key that unlocks the entire equilibrium constant calculation.

THE CLASSIC MODEL SYSTEM: Fe^{3+} AND SCN^-

The most widely taught laboratory experiment for this topic uses the reaction between iron(III) ions and thiocyanate ions to form the blood-red iron(III) thiocyanate complex:



This reaction is ideal for colorimetric analysis because:

- The product, FeSCN^{2+} , is intensely colored (deep red-orange), while the reactants Fe^{3+} (pale yellow) and SCN^- (colorless) absorb very little visible light at the wavelength of maximum absorption for the complex (around 447–470 nm).
- The reaction is reversible and reaches equilibrium quickly at room temperature.
- The stoichiometry is 1:1, simplifying calculations.

Other possible systems include the copper(II)-ammonia complex or the iron(III)-salicylate complex, but the $\text{Fe}^{3+}/\text{SCN}^-$ system remains the standard for introductory and advanced courses alike.

EXPERIMENTAL METHODOLOGY FOR EQUILIBRIUM CONSTANT DETERMINATION

Materials and Instrumentation

To perform a colorimetric equilibrium constant determination, you will need:

- A spectrophotometer or colorimeter capable of measuring absorbance in the visible range

(400–600 nm)

- Standard volumetric glassware (volumetric flasks, pipettes, burettes)
- Solutions of known concentrations: typically 0.002 M $\text{Fe}(\text{NO}_3)_3$ in 1 M HNO_3 (the nitric acid suppresses hydrolysis of Fe^{3+}) and 0.002 M KSCN (or NaSCN)
- Deionized water for dilutions
- Cuvettes (usually 1 cm path length)

Step 1: Preparation of Standard Solutions for the Calibration Curve

A calibration curve is constructed using solutions that contain only the product, FeSCN^{2+} , at known concentrations. However, because the equilibrium constant is not extremely large, it is difficult to prepare pure product solutions. Instead, we use the "reactant excess" method: prepare solutions where SCN^- is the limiting reagent and Fe^{3+} is in large excess. Under these conditions, the equilibrium shifts far to the right, and essentially all SCN^- is converted to FeSCN^{2+} . Thus, the concentration of FeSCN^{2+} can be assumed equal to the initial concentration of SCN^- .

For example, prepare six standard solutions in 25 mL volumetric flasks: add 5.00 mL of 0.20 M $\text{Fe}(\text{NO}_3)_3$ (excess) and varying volumes (e.g., 0.50, 1.00, 1.50, 2.00, 2.50, 3.00 mL) of 0.002 M KSCN,

then dilute to the mark with 1 M HNO_3 . The final concentration of FeSCN^{2+} in each flask equals (volume KSCN $\times$ 0.002 M) / total volume.

Step 2: Measuring Absorbance and Building the Calibration Curve

Set the spectrophotometer to the wavelength of maximum absorbance for FeSCN^{2+} (typically 447 nm). Zero the instrument using a blank solution (1 M HNO_3). Measure the absorbance of each standard solution. Plot absorbance (y-axis) versus concentration of FeSCN^{2+} (x-axis). The plot should be linear through the origin, confirming the Beer-Lambert Law. Determine the slope of the best-fit line, which equals $\epsilon \times b$ (with $b = 1$ cm, slope = ϵ).

Step 3: Preparing Equilibrium Mixtures

Now prepare a series of equilibrium mixtures where Fe^{3+} and SCN^- are both at low, comparable initial concentrations. For example, in separate 25 mL flasks, mix volumes of 0.002 M $\text{Fe}(\text{NO}_3)_3$ and 0.002 M KSCN such that the initial concentrations are roughly in the range of $1\text{--}4 \times 10^{-4}$ M. Do not add excess Fe^{3+} . Dilute each to the mark with 1 M HNO_3 .

Allow the solutions to stand for a few minutes to ensure equilibrium is reached.

Step 4: Measuring Equilibrium Absorbance

Measure the absorbance of each equilibrium mixture at the same wavelength used for the calibration curve. Record the absorbance values precisely.

DATA ANALYSIS: CALCULATING THE EQUILIBRIUM CONSTANT

Determining the Equilibrium Concentration of the Colored Product

Using the calibration curve (or the Beer-Lambert equation), convert the measured absorbance of each equilibrium mixture to the concentration of FeSCN^{2+} at equilibrium:

$$[\text{FeSCN}^{2+}]_{\text{eq}} = \text{Absorbance} / \text{slope}$$

Calculating

Equilibrium Concentrations of Reactants

From the stoichiometry of the reaction (1:1:1), the amount of Fe^{3+} and SCN^- consumed equals the amount of FeSCN^{2+} formed. Therefore:

$$[\text{Fe}^{3+}]_{\text{eq}} = [\text{Fe}^{3+}]_{\text{initial}} - [\text{FeSCN}^{2+}]_{\text{eq}}$$

$$[\text{SCN}^-]_{\text{eq}} = [\text{SCN}^-]_{\text{initial}} - [\text{FeSCN}^{2+}]_{\text{eq}}$$

Where the initial concentrations are calculated from the volumes and stock concentrations used in each equilibrium mixture, accounting for dilution in the 25 mL flask.

Computing the Equilibrium Constant

For each equilibrium mixture, calculate K_c using:

$$K_c = [\text{FeSCN}^{2+}]_{\text{eq}} / ([\text{Fe}^{3+}]_{\text{eq}} \times [\text{SCN}^-]_{\text{eq}})$$

Because the reaction is 1:1, the units of K_c will be L mol^{-1} . Compute the average K_c value from all mixtures. At room temperature (25 °C) and in 1 M HNO_3 , the expected K_c for the FeSCN^{2+} complex is typically around 200–300 L mol^{-1} , though it varies with ionic strength and acid concentration.

Error Analysis and Refinements

Replicate measurements and multiple equilibrium mixtures help reduce random error. Systematic errors may arise from:

- Incomplete conversion in the standard solutions (if Fe^{3+} excess is insufficient). Using at least 50-fold excess of Fe^{3+} ensures >99% conversion of SCN^- .
- Absorbance contributions from Fe^{3+} itself. This is minimized by using a blank that contains the same concentration of Fe^{3+} as in the samples, or by subtracting a small baseline absorbance.
- Variations in temperature and ionic strength. Performing all measurements at the same temperature and maintaining high and constant ionic strength (via 1 M HNO_3) improves precision.

PRACTICAL CONSIDERATIONS AND TROUBLESHOOTING

Choosing the Right Wavelength

If using a spectrophotometer, first scan the absorbance spectrum of a representative FeSCN^{2+} solution to identify λ_{max} . For a colorimeter, select a filter that offers maximum absorbance. Avoid wavelengths where the blank or reactants absorb significantly.

Handling Hydrolysis and Side Reactions

Fe^{3+} ions hydrolyze in water to form FeOH^{2+} and other species, which can interfere. The use of 1 M HNO_3 maintains the solution strongly acidic, suppressing hydrolysis and ensuring that Fe^{3+} remains as the hexaaqua ion, which does not absorb significantly at the analysis wavelength.

Instrument Care

Always use clean, scratch-free cuvettes. Handle cuvettes by the frosted sides to avoid fingerprints on the optical windows. Rinse cuvettes with the solution to be measured before filling, to eliminate dilution errors.

VARIATIONS AND EXTENSIONS OF THE METHOD

The same colorimetric approach can be applied to other equilibrium systems. For example, the formation of the copper-ammonia complex $\text{Cu}(\text{NH}_3)_4^{2+}$ (deep blue) can be used to determine the stepwise formation constants. In that case, the equilibrium is more complex, and multiple species may absorb, requiring careful selection of wavelength or use of a diode-array spectrophotometer.

For reactions with 1:2 or other stoichiometries, the stoichiometric relationships change accordingly. The fundamental principle remains: measure the concentration of one species colorimetrically at equilibrium, then use initial concentrations and the balanced

equation to calculate all equilibrium concentrations and thus K_c .

Advanced variants include the use of **spectrophotometric titration**, where absorbance is recorded as a function of

added titrant, and the equilibrium constant is extracted from the inflection point or by nonlinear regression of the entire titration curve. This is particularly useful when the absorbing species is a reactant or product in a weak acid-base